
User's Guide RIGOL User's Guide RIGOL



NFP-3 Near Field Probe


Product Overview

NFP-3 is used with **RIGOL** DSA series spectrum analyzer for the EMI tests of electronic products. It can be used to test the magnetic field strength and magnetic field coupling channels on the surface of the electronic components as well as the magnetic field environment near the electronic module so as to quickly locate the interference source. NFP-3 includes four models (NFP-3-P1, NFP-3-P2, NFP-3-P3 and NFP-3-P4).



Measurement Connections

The connection mode of NFP-3 and spectrum analyzer is as shown in the figure below.



- Connect the spectrum analyzer

Connect the SMB (M) terminal of NFP-3 and the BNC (F) terminal of the N-BNC adaptor respectively via the BNC-SMB RF cable; connect the N (M) terminal of the N-BNC adaptor to the RF input terminal of the spectrum analyzer.

Connect the device under test

NFP-3 is used to perform short-distance noncontact measurement on the device under test. Pay attention to the direction of the probe during measuring.

Typical Applications

- Locate the EMI radiation interference source.
- Determine the frequency and relative strength of the spectral component of the interference source.

Performance Specifications

Frequency	
Frequency Range	30 MHz to 3 GHz

Terminal Type	
Terminal Type	SMB (M)
Adaptor	N (M)-BNC (F)
RF Cable	BNC (M)-SMB (F), 1000 mm
Terminal and Adaptor Impedance	50 Ω

General Specifications	
Dimensions	260 mm \times 190 mm \times 30 mm (with package)
Weight	0.425 kg (with package)
Operating Temperature	0°C to 50°C
Storage Temperature	-20℃ to 70℃

Aug. 2015

Model	Description	
NFP-3-P1	Near field probe for magnetic field measurements. The test range	
THE THE PROPERTY OF THE PROPER	is within 10 cm. It is used to locate the leakage field. P1 -10 -20 -40 -50 -50 -50 -50 -50 -60 -50 -60 -70 -70 -70 -70 -70 -70 -7	
NED 2 D2		
NFP-3-P2	Near field probe for magnetic field measurements. The test range is within 3 cm. It is used to accurately test the leakage field.	
H	-10	
	-60 0 500 1000 1500 2000 2500 3000 Frequency (MHz)	

