GOOD WILL INSTRUMENT CO., LTD. N0. 7-1, Jhongsing Road, Tucheng Dist., New Taipei City, 236, Taiwan T (886) 2 2268-0389 F (886)2 2268-0639 www.gwinstek.com

**G<u><u><u></u></u>INSTEK**</u>

## **APS-7000 Specifications**

The specifications apply when the APS-7000 is powered on for at least 30 minutes under +20°C~+30°C.



Model		APS-7050	APS-7100	APS-7200	APS-7300	
Power Rating		500VA	1000VA	2000VA	3000VA	
Output Voltage		0 ~ 155Vrms, 0 ~ 310Vrms				
Output Frequency		45.00 ~ 500.0 Hz				
Max. Current(rms) *1	0 ~ 155 Vrms	4.2A	8.4A	16.8A	25.2A	
	0 ~ 310 Vrms	2.1A	4.2A	8.4A	12.6A	
Max. Current (peak)	0 ~ 155 Vrms	16.8A	33.6A	67.2A	100.8A	
	0 ~ 310 Vrms	8.4A	16.8A	33.6A	50.4A	
OPT. APS-003(rms) *6	0 ~ 600 Vrms	1.05A	2.1A	4.2A	6.3A	
OPT. APS-003(peak)	0 ~ 600 Vrms	4.2A	8.4A	16.8A	25.2A	
Total Harmonic Distortion (THD) *2		≤ 0.5% at 45 ~ 500.0Hz (Resistive Load)				
Crest factor		≤ 4				
Line regulation		0.1% (% of full scale)				
Load regulation		0.3% (% of full scale)				
Response time		< 100µs				
Reverse current		30% of Maximum Output RMS Current (Continue) 100% of Maximum Output RMS Current (Within 3 minutes)				
Setting						
Voltage	Range	0 ~ 155Vrms, 0 ~ 310Vrms, Auto				
	Resolution	0.01V at 0.00 ~ 99.99Vrms 0.1V at 100.0 ~ 310.0Vrms				
	Accuracy	± (0.5% of setting + 2 counts)				

<u>G</u>UINSTEK

## GOOD WILL INSTRUMENT CO., LTD.

N0. 7-1, Jhongsing Road, Tucheng Dist., New Taipei City, 236, Taiwan T (886) 2 2268-0389 F (886)2 2268-0639 www.gwinstek.com

Frequency	Range	45 ~ 500 Hz			
	Resolution	0.01Hz at 45.00 ~ 99.99Hz 0.1Hz at 100.0 ~ 500.0Hz			
	Accuracy	± 0.02% of Setting			
Power on/off phase angle	Range	0~.	0 ~ 359°		
	Resolution	1°			
	Accuracy	± 1° (Only 45 ~ 65Hz)			
Measurement *3					
Voltage (rms)	Range	0.20 ~ 38.75Vrms 38.76 ~ 77.50Vrms 77.51 ~ 155.0Vrms 155.1 ~ 310.0Vrms			
	Resolution	0.01V at 0.00 ~ 99.99Vrms 0.1V at 100.0 ~ 310.0Vrms			
	Accuracy *4	± (0.5% of reading + 2 counts)			
Frequency	Range	45 ~ 500Hz			
	Resolution	0.01Hz at 45.00 ~ 99.99Hz 0.1Hz at 100.0 ~ 500.0Hz			
	Accuracy *5	± 0.1Hz			
Current (rms)	Range	2.00 ~ 70.00mA 60.0 ~ 350.0mA 0.300 ~ 3.500A 3.00 ~ 17.50A	0.200 ~ 3.500A 3.00 ~ 35.00A		
	Resolution	0.01mA 0.1mA 0.001A 0.01A	0.001A 0.01A		
	Accuracy	<ul> <li>± (0.6% of reading + 5 counts) , 2.00 ~ 350.0mA</li> <li>± (0.5% of reading + 5 counts) , 0.300 ~ 3.500A</li> <li>± (0.5% of reading + 3 counts) , 3.000 ~ 17.50A</li> </ul>	± (0.5% of reading + 5 counts) , 0.200 ~ 3.500A ± (0.5% of reading + 3 counts) , 3.00 ~ 35.00A		
Current (peak) Power (W)	Range	0.0 ~ 70.0A	0.0~140.0A		
	Resolution	0.1A	0.1A		
	Accuracy	± (1% of reading + 1 count)	± (1% of reading + 1 count)		
	Resolution	0.01W 0.1W 1W	0.1W 1W		
	Accuracy	± (0.6% of reading + 5 counts) , 0.20 ~ 99.99W ± (0.6% of reading + 5 counts) , 100.0 ~ 999.9W	± (0.6% of reading + 5counts) , 0.2 ~ 999.9W ± (0.6% of reading + 2counts) , 1000 ~ 9999W		

**G<u><u><u></u></u>INSTEK**</u>

## GOOD WILL INSTRUMENT CO., LTD.

N0. 7-1, Jhongsing Road, Tucheng Dist., New Taipei City, 236, Taiwan T (886) 2 2268-0389 F (886)2 2268-0639 www.gwinstek.com

			+ 2 counts) , 1000 ~ 99W				
Apparent Power (VA)	Resolution	0.1	1VA LVA VA	0.1 1V			
	Accuracy	99.9 ± (1% of reading + 999 ± (1% of reading +	- 7 counts) , 0.20 ~ 99VA 7 counts) , 100.0 ~ .9VA • 5 counts) , 1000 ~ 9VA	± (1% of reading + 999. ± (1% of reading + 999!	9VA 5 counts) , 1000 ~		
Power Factor	Range	0.000~ 1.000					
	Resolution	0.001					
	Accuracy	±2% reading + 2 counts					
General							
Remote output signal		Pass, Fail,	Test-in Process, Trigg	er in, Trigger out, OU	T ON/OFF		
Sync output signal			Output Signal 10 V, BNC Type				
Number of Preset			10 (0~9 numeric keys)				
Protection			OCP, OPP, OTP and Alarm				
Trigger Out		Maximum low level output = 0.8V Minimum high level output = 2V Maximum source current = 8mA.					
Trigger In		Maximum low level input voltage = 0.8V Minimum high level input voltage = 2.0V Maximum sink current = 8mA.					
SEQUENCE / SIMULATIO	ON FUNCTION	1					
Number of Memories			10 (0~9 nu	meric keys)			
Number of Steps		255 max. (for 1 sequence)					
Step Time Setting Range	*9	0.01 ~ 999.99 s					
Operation within Step		Constant, Keep, Linear Sweep					
Parameters		Output Range, Frequency, Waveform (sine wave only), On Phase, Off Phase, Term Jump Count (0 ~255), Jump-to, Branch 1, Branch 2, Trigger Output					
Sequence Control		Start, Stop, Hold, Continue, Branch 1, Branch 2					
AC Input							
Phase		Single Phase					
Input Voltage		115/230Vac ± 15% 230Vac ± 15%			: ± 15%		
Input Frequency		50/60Hz					
Max. Current		16A/8A	32A/16A	32A	50A		
Power Factor		0.7 Тур.					

## GOOD WILL INSTRUMENT CO., LTD.

**G<u><u><u></u></u>INSTEK</u>** 

N0. 7-1, Jhongsing Road, Tucheng Dist., New Taipei City, 236, Taiwan T (886) 2 2268-0389 F (886)2 2268-0639 www.gwinstek.com

Environmental Conditions						
Operating temperature range		0 ~ +40 °C				
Storage temperature range		-10 ~ +70 °C				
Operating humidity range		20 ~ 80% RH (no condensation)				
Storage humidity range		80% RH or less (no condensation)				
Interface						
Standard		USB Host, LAN		USB Host, USB CDC, LAN		
Optional		GPIB (APS-001) RS232 / USB CDC (APS-002)		GPIB (APS-001) RS232 (APS-007)		
Dimensions & Weight						
Dimensions (mm)	W	430	430	430	430	
	н	88	88	312	400	
	D	400	560	650	650	
Weight		24kg	38kg	90kg	128kg	

Product specifications are subject to change without notice.

The Specifications are not suit for ARB mode.

- \*1 Maximum output current at working voltage 120Vrms, 240Vrms
- \*2 45 to 500Hz, 10% or higher of the rated output voltage, the maximum current or lower
- \*3 All of measurement accuracy is at 23±5°C
- \*4 In the case of 15 to 155V, 30 to 310V, sine wave, no load
- \*5 For an output voltage of 75 V to 155 V / 150 V to 310 V, sine wave, no load
- \*6 Maximum output current at working voltage 480Vrms
- \*9 Step Time does not guarantee specification.