

High Voltage Potential Insures Your Potential



GPT-15012

Electrical Safety Analyzer

- DC 12kV / IR 5kV
- Comply with IEC 61010-2-034
- 7" TFT LCD
- Manual / Auto Mode
- RMS Current Measurement
- Zero Crossing Turn-on Operation
- Controllable Ramp-up & Ramp-down Time
- Capacitive Load Testing Capability up to 47 μ F
- Statistics & Analysis Function
- Sweep Function for DUT Characteristic Analysis
- Internal Storage & USB Storage available
- Rear panel output available
- Interface : RS-232C, USB host/device, Signal I/O and option GPIB or LAN



DC WITHSTANDING	
Output-Voltage Range	0.050kV~12.00kV
Output-Voltage Resolution	1V
Output-Voltage Accuracy	± (1% of setting + 5V) [no load]
Maximum Rated Load	120W (12kV/10mA)
Maximum Rated Current	10mA (0.5kV< V ≤12kV); 2mA (0.05kV≤ V ≤0.5kV)
Voltage Regulation	± (1% + 5V) [maximum rated load → no load]
Voltmeter Accuracy	± (1% of reading + 5V)
Current Measurement Range	1μA~10.00mA
Current Best Resolution	0.1μA / 1μA / 10μA
Current Measurement Accuracy	±(1.5% of reading + 3μA) when I Reading < 1mA ±(1.5% of reading + 30μA) when I Reading ≥ 1mA
Current Offset	5μA Maximum
Window Comparator Method	Yes
ARC Detect	Yes
RAMP UP (Rise Time)	0.1s~999.9s
RAMP DOWN (Fall Time)	0.0s~999.9s
TIMER (Test Time)	OFF, 0.3s~999.9s
TIMER Accuracy	±(100ppm + 20ms)
WAIT TIME	0.0s~999.9s
GND	ON/OFF
INSULATION RESISTANCE	
Output Voltage	50V~5000V dc
Output-Voltage Resolution	50V
Output-Voltage Accuracy	± (1% of setting + 5V) [no load]
Resistance Display	
Resistance Measurement	
Test Voltage	Measurement Range / Accuracy
50V≤V≤450V	0.1MΩ~1MΩ : ±(5% of reading + 3 count) 1.1MΩ~50MΩ : ±(5% of reading + 1 count) 50.1MΩ~2GΩ : ±(10% of reading + 1 count)
500V≤V≤1200V	0.1MΩ~1MΩ : ±(5% of reading + 3 count) 1.1MΩ~500MΩ : ±(5% of reading + 1 count) 500.1MΩ~9.999GΩ : ±(10% of reading + 1 count) 10GΩ~50GΩ : ±(20% of reading + 1 count)*
1250V≤V≤5000V	0.1MΩ~1MΩ : ±(5% of reading + 3 count) 1.1MΩ~500MΩ : ±(5% of reading + 1 count) 500.1MΩ~9.999GΩ : ±(10% of reading + 1 count) 10GΩ~50GΩ : ±(15% of reading + 1 count)*
Voltage Regulation	± (1% + 5V) [maximum rated load → no load]
Voltmeter Accuracy	± (1% of reading + 5V)
Short-Circuit Current	10mA max.
Output Impedance	2kΩ
Window Comparator Method	Yes
RAMP UP (Rise Time)	0.1s~999.9s
RAMP DOWN (Fall Time)	0.0s~999.9s
TIMER (Test Time)	OFF, 0.3s~999.9s
TIMER Accuracy	±(100ppm + 20ms)
WAIT TIME	0.0s~999.9s
GND	ON/OFF
MEMORY	
Single Step Memory	MANU : 100 blocks
Automatic Testing Memory	AUTO : 100 blocks, manu per auto : 10
INTERFACE	
Front Panel	USB host, REMOTE
Rear Panel	RS-232C, USB device, Signal I/O, Rear Output
Option	GPIO, LAN
DISPLAY	
	7" color LCD
POWER SOURCE	
	AC 100V~240V ± 10%, 50Hz/60Hz; Power consumption : 400VA max.
DIMENSION & WEIGHT	
	380(W) x 148(H) x 492(D) mm; Approx. 9.8kg

*: Detail refers to user manual

Specifications subject to change without notice.

Global Headquarters

GOOD WILL INSTRUMENT CO., LTD.

No.7-1, Jhongsing Road, Tucheng Dist., New Taipei City 236, Taiwan
T +886-2-2268-0389 F +886-2-2268-0639
E-mail: marketing@goodwill.com.tw

China Subsidiary

GOOD WILL INSTRUMENT (SUZHOU) CO., LTD.

No. 521, Zhujiang Road, Snd, Suzhou Jiangsu 215011 China
T +86-512-6661-7177 F +86-512-6661-7277

Malaysia Subsidiary

GOOD WILL INSTRUMENT (SEA) SDN. BHD.

No. 1-3-18, Elit Avenue, Jalan Mayang Pasir 3,
11950 Bayan Baru, Penang, Malaysia
T +604-6111122 F +604-6115225

Europe Subsidiary

GOOD WILL INSTRUMENT EURO B.V.

De Run 5427A, 5504DG Veldhoven, THE NETHERLANDS
T +31(0)40-2557790 F +31(0)40-2541194

U.S.A. Subsidiary

INSTEK AMERICA CORP.

5198 Brooks Street Montclair, CA 91763, U.S.A.
T +1-909-399-3535 F +1-909-399-0819

Japan Subsidiary

TEXIO TECHNOLOGY CORPORATION.

7F Towa Fudosan Shin Yokohama Bldg., 2-18-13 Shin
Yokohama, Kohoku-ku, Yokohama, Kanagawa,
222-0033 Japan
T +81-45-620-2305 F +81-45-534-7181

Korea Subsidiary

GOOD WILL INSTRUMENT KOREA CO., LTD.

Room No.503, Gyeonginro 775 (Mullae-Dong 3Ga,
Ace Hightech-City B/D 1Dong), Yeongduengpo-Gu,
Seoul 150093, Korea.
T +82-2-3439-2205 F +82-2-3439-2207

India Subsidiary

GW INSTEK INDIA LLP.

No.2707/B&C, 1st Floor UNNATHI Building,
E-Block, Sahakara Nagar, Bengaluru-560 092, India
T +91-80-6811-0600 F +91-80-6811-0626

GW INSTEK
Simply Reliable



Website



Facebook



LinkedIn